

BIOHIT HELICOBACTER PYLORI UFT300

quick test with conical tube

QUICK GUIDE

It is recommended that the biopsy specimen is tested immediately after collection. Do not let the specimen dry before testing. If there is visible blood or bile on the specimen, dab the biopsy briefly on a sterile gauze.

1.
 - Write the ID of the patient on the tube label (provided on separate sheets).
 - Place one or more biopsies from the same patient into the tube.



2.
 - Add one drop of the reagent solution into the test tube - just the minimum amount required to cover the biopsy (fig. 1). Keep the bottle upright while dropping (fig. 2).
 - **NOTE!** Be careful not to add too much of the reagent solution to the biopsy – just add the (minimum) amount required to cover the biopsy.
 - Put the cap back on the tube.
 - Shake the tube by swirling it horizontally on a flat surface for 5 seconds.

Figure 1



Figure 2



3.
 - Incubate for 5 minutes.
 - Remove the cap and interpret the result by looking into the tube from the top.

Negative result: Color of the test solution remains yellow.
Positive result: Color of the test solution turns from yellow to magenta.

Do not read the color after 5 min.



Negative



Positive

